

Declarations



Hereby we declare

Xylem Analytics Germany Sales GmbH & Co. KG, ebro
Peringerstraße 10
85055 Ingolstadt, Germany
Phone: +49 841 95478-0
Fax: +49 841 95478-80

that the following product

Product type:	Data logger
Type designation:	EBI 25-T / -TE / -TX / -TH, EBI 20-T1 / -TE1 / -TF / -TH1, EBI 300, EBI 310, EBI 330

is in compliance with the essential requirements and other relevant provisions of Directive 37/2005 EC.

The following harmonized standards have been used:

- **Tests, performance, suitability: EN 12830**
- **Periodic verification and calibration: EN 13486**

Eckehard Peschel, Site Leader

Hereby we declare

Xylem Analytics Germany Sales GmbH & Co. KG, ebro
Peringerstraße 10
85055 Ingolstadt, Germany
Phone: +49 841 95478-0
Fax: +49 841 95478-80

that the following product

Product type:	Thermometer
Type designation:	TLC 700, TLC 750i, TLC 750 BT, TLC 750 NFC, TLC 1598, TFX 410, TFX 410-1, TFX 420, TFX 422 C, TFX 430, TTX 110, TTX 2x0, TFE 510, TMX 3x0, TMX 4x0

is in compliance with the essential requirements and other relevant provisions of Directive 37/2005 EC.

The following harmonized standards have been used:

- **Tests, performance, suitability: EN 13485**
- **Periodic verification and calibration: EN 13486**

Eckehard Peschel, Site Leader